

**Certificate of Conformance
European Community
Council Directive 2004/108/EC**

Date of Issue: November, 2008
Issued By: Retlif Testing Laboratories
795 Marconi Avenue
Ronkonkoma, NY 11779

Issued To: Climatronics Corp.
140 Wilbur Place
Bohemia, NY 11716

Reference: Retlif Report Number R-12669

Retlif Testing Laboratories hereby acknowledges that compliance testing in accordance with the below listed standards was performed on a representative sample of the equipment listed below. Retlif Testing Laboratories further acknowledges that the test sample listed below was found to be in compliance with these standards.

This certificate is hereby issued to the above named grantee and is valid only for the equipment identified below.

Manufacturer: Climatronics Corp.
140 Wilbur Place
Bohemia, NY 11716

Equipment Tested: Barometric Pressure Sensor

Part Number: 102663/092

Serial Number: 42288

Brand Name: Climatronics Corp./Met One Instruments

Product Type: Measurement, Control Equipment and Laboratory Use

- Note(s):
- 1) See attached report R-12669 for details and/or conditions pertaining to this certificate.
 - 2) Conforms to the emissions requirements of EN 61326-1:2006; Clause 7.2
CISPR 11:1Edition 4 2003 Class A, Radiated Emissions, 30 MHz to 1 GHz
 - 3) Conforms to the immunity requirements of EN 61326:2006-1:2006; Table 2
IEC 61000-4-2:2001 Electrostatic Discharge
IEC 61000-4-3:2002 Radiated Immunity
IEC 61000-4-4:2004 EFT/Burst, Power and I/O Leads
IEC 61000-4-5:2001 Surge Immunity, Power Leads
IEC 61000-4-6:2003 Conducted Immunity, Power and I/O Leads